

Nehru Arts and Science College Kanhangad

Tender Reference No.: LTEN/NASCFIST/2023/1

Dated: 30.03.2023

SPECIFICATIONS

SL No	Item	Specification/useful for
1	Spray Pyrolysis equipment and accessories	Hot Plate Size: 150 mm x150 mm, maximum temperature: 500degree Celsius or more, PID controller for Hot Plate Temperature, syringe pump for dispensing with, Automatic XY movement spray mechanism, dispenser, substrate heater, compressor, stepper motor controller, exhaust system in-house, standard air automizing nozzle and software.

Note: The specifications for the items quoted should be explicitly clear and all available technical literature, catalogues and other data in support of the specification of the items should be furnished along with the tenders.

Dr. K V Murali, Principal
(Name & designation of Purchasing Officer)

Station and date: Padnekad, 30-03-2023.